



**Doctors For You, India
(DFY India)**

Bid Document

**Doctors For You Request for Proposal (RFP) For Supply of Health Centre Medical Items at
C/O Manik Chandra Hatibaruah, Holongapar Gohain Gaon, Lahdoigarh, Pin No: 785700,
District: Jorhat.**

Ref. No.: DFY/26-27/10/05/TEN01

Doctors For You

Flat No - 201/203, Building Number – 22,
Lallubhai Compound, Mankhurd,
Mumbai-400043, Maharashtra, INDIA
Email: purchase.department@doctorsforyou.org
Website: www.doctorsforyou.org



Bid Document

Doctors For You Request for Proposal (RFP) For Supply of Health Centre Medical Items at C/O Manik Chandra Hatibaruah, Holongapar Gohain Gaon, Lahdoigarh, Pin No: 785700, District: Jorhat.

Key Bidding Information

Bid Ref No.	DFY/26-27/10/05/TEN01
Name of the Project	Assam Flood Response 2026
Date of Commencement of Download of Bidding Documents	10 th October 2026 www.doctorsforyou.org
Last Date and Time for Receipt of Request for Clarifications	15 th October 2026 till 04:00 PM purchase.department@doctorsforyou.org
Last Date & Time for Submission of Bids	16 th October 2026 till 01:00 PM purchase.department@doctorsforyou.org
Date & Time for Opening of Bids	17 th October 2026 at 01:00 PM



Doctors For You invites the bids from agencies/vendor for Supply of Health Centre Medical Items at C/O Manik Chandra Hatibaruah, Holongapar Gohain Gaon, Lahdoigarh, Pin No: 785700, District: Jorhat.

1. Doctors For You (DFY), is a registered society, registered under the Societies Registration Act 1860 Section 21 having registration no. F-56886(Mum). DFY is a pan India humanitarian organization with international presence and is working in various disaster hit zones since last 19 years. DFY focuses on providing medical care to the vulnerable communities during crisis and non-crisis situation, emergency medical aid to people affected by natural disaster, conflicts and epidemics. We are also committed to reducing disaster risk to human society by delivering trainings and capacity development in emergency preparedness and response. The work of DFY is guided by humanitarian principles of humanity, impartiality, and neutrality. It offers services and assistance to people based on need, irrespective of race, class, caste, religion and gender.

Doctors For You invites the bids from agencies/vendors for the Supply of Health Centre Medical Items at C/O Manik Chandra Hatibaruah, Holongapar Gohain Gaon, Lahdoigarh, Pin No: 785700, District: Jorhat

2. The detail of requirement has been given in Section III.
3. DFY hereby invites bids (through email or hardcopy at email id (purchase.department@doctorsforyou.org) or address (Doctors For You' Flat no. G6/312 Pocket – D, Ganga Apartment Vasant Kunj, Delhi – 110070) from eligible and qualified Bidders for the Supply of Health Centre Medical Items at C/O Manik Chandra Hatibaruah, Holongapar Gohain Gaon, Lahdoigarh, Pin No: 785700, District: Jorhat.
4. Bidding will be conducted through 'Tender Enquiry' method and procedures as per DFY's procurement policy.
5. Interested agencies/vendors provide their bid price on a per unit basis. The final payment will be based on the deliveries as per the timeline given.
6. Vendors are requested to submit their detailed specifications and financial quotations, including all applicable taxes, transportation, installation, commissioning, and other associated costs.
7. Doctors For You may coordinate with the vendor for any clarifications regarding the submitted quotations, if required, within 4 days from the tender opening date.
8. Bidders can bid for one or multiple items at a time.
9. The Bid Document can be freely downloaded from the website www.doctorsforyou.org Starting from **10th October 2026**. Bidders shall be solely responsible for checking the above website for any addendum/amendment issued subsequent to publication of this Invitation for Bids and take the same into consideration while preparing and submitting their bids.
10. **Bids must be submitted on or before 01:00 PM on 16th October 2026 by Bidders** through email or hardcopy through courier
11. **The Bids will be opened on 17th October 2026 at 01:00 PM.**

SECTION – I INSTRUCTIONS TO BIDDERS

A. INTRODUCTION	
1. Documents Establishing conformity of Goods and Services to Bidding Documents	1.1 The Bidder shall furnish as part of its Bid the documentary evidence that the Goods/services conform to the technical specifications and standard specified in Scope of work. The documentary evidence of conformity of the goods and services to the Bidding Documents may be in the form of brochure / technical literature. 1.2 GST and Income Tax registration certificates.
2. Cost of Bidding	2.1 The Bidder shall bear all costs associated with the preparation and submission of its bid, and the Purchaser will in no case be responsible or liable for those costs, regardless of the conduct or outcome of the bidding process.
B. THE BIDDING DOCUMENTS	
3. Clarification of Bidding Documents	3.1 A prospective bidder requiring any clarification regarding specification, required services and conditions of contract/agreement, etc. given in the document may submit a written request for clarifications to by email up to 16:00 hours on 15 th October 2026 at email id: purchase.department@doctorsforyou.org
C. PREPARATION OF BIDS	
4. Documents Constituting the Bid	4.1 The Bid shall comprise two parts submitted simultaneously, one part containing the Bid and the other part containing the Financial Bid. 4.2 At first, the Technical Bids shall be opened at the time and date specified in the Table – Key Information given at page no-2 4.3 The Technical Bids shall be evaluated first by the Purchaser. The Technical Bids which do not conform to the specified requirements will be rejected as technically non-responsive/disqualified Bids. 4.4 Financial Bids of technically qualified Bids shall be opened by the Purchaser at a later date and time communicated to all technically qualified bidders. The Financial Bids shall be evaluated for all requirements and the Contract shall be awarded to the Bidder who has been determined to be the lowest evaluated substantially responsive Bidder for full requirement. 4.5 The Financial Bid submitted by the Bidder shall comprise the duly filled-in Price Schedule, in accordance with the forms indicated in Section IV.

5. Bid Prices	5.1 The Bidder shall indicate on the Price Schedule as per the format of Price Schedule given in Section IV. 5.2 The rate quoted should be both in words and figures. No figure or word should be overwritten. Correction if any should be rewritten under the full signature of the person signing the bid. 5.3 The prices quoted by the bidder should be on a firm and fixed basis during the performance of the contract, except for duties and taxes on finished goods. A bid submitted with an adjustable price quotation will be treated as non-responsive and will be rejected.
6. Currencies of Bid	6.1 Prices shall be quoted in Indian Rupees only.
7. Period of Validity of Bids	7.1 Bids shall remain valid for the period of 30 days after the date of bid submission specified in the document. 7.2 In exceptional circumstances, prior to the expiry of the original bid validity period, the Purchaser may request that the Bidders extend the period of validity for a specified additional period. The request and the responses there to shall be made in writing.
8. Alternative Proposals by Bidders	8.1 Alternative bids shall not be accepted. The bidder should not submit more than one bid.
9. Format and Signing of Bid	9.1 The bids consisting of the documents listed in this document shall be typed or written in indelible ink and shall be signed by the Bidder or a person or persons duly authorized to bind the Bidder to the Contract. 9.2 Any interlineations, erasures, or overwriting to correct errors made by the Bidder should be initiated by the person or persons signing the bid.
D. Submission of Bids	
10. Sealing and Marking of Bids	10.1 Bidders need to submit their bids electronically through email ID : purchase.department@doctorsforyou.org or can send hard copy through courier at 'Doctors For You' Flat no. G6/312 Pocket – D, Ganga Apartment Vasant Kunj, Delhi – 110070.

11. Deadline for Submission of Bids	11.1 Bids must be received by the Purchaser no later than the time and date as specified in the Table – Key Information on page no. 2 of the Bid Document. 11.2 In event of the specified date for the submission of Bids being declared a holiday for the Purchaser, the Bids will be received up to the appointed time on the next working day. 11.3 The Purchaser may, at its discretion, extend the deadline for the submission of bids by amending the Bidding Documents in accordance with ITB Sub-Para 7.2, in which case all rights and obligations of the Purchaser and Bidders previously subject to the deadline will thereafter be subject to the deadline as extended.
12. Modification and Withdrawal of Bids	12.1 The bids can't be withdrawn or altered after the due date and time for the submission of bids.
E. OPENING AND EVALUATION OF BIDS	
13. Bid Opening	13.1 The financial bids shall be opened by the Bid Opening Committee of DFY's through an online platform or in the presence of committee members
14. Clarification of Bids	14.1 During the evaluation of the bids, the Purchaser may, at its discretion, ask the Bidder for a clarification of its bid. The request for clarification and the response shall be in writing, and no change in the prices or substance of the bid shall be sought, offered, or permitted, except to correct arithmetic errors identified by the Purchaser in the evaluation of the bids.
15. Examination of Bids and Determination of Responsiveness	15.1 The Purchaser's determination of the responsiveness of a Technical Proposal is to be based on the contents of the Technical Proposal itself. 15.2 The Purchaser will examine the technical bids to determine whether they are complete, whether the documents have been properly signed, and whether the bids are generally in order. 15.3 The Purchaser may waive any minor informality, nonconformity, or irregularity in a bid that does not constitute a material deviation, provided such waiver does not prejudice or affect the relative ranking of any Bidder.

16. Evaluation and Comparison of Bid	<u>Preliminary Evaluation of Bids</u>The Purchaser shall examine the Technical Proposal to confirm that all documents and technical documentation requested in RFP have been provided, and to determine the completeness of each document submitted. 16.1 The Purchaser shall confirm that the required documents and information have been provided in the Technical Proposal. 16.2 Likewise, following the opening of Price Proposals, the Purchaser shall examine the Price Proposals to confirm that all documents and financial documentation have been provided and to determine the completeness of each document submitted. <u>Evaluation of Price Bids</u> 16.3 The Purchaser shall evaluate Price Proposals of each Bid for which the Technical Proposal has been determined to be responsive. 16.4 The Purchaser's evaluation of a bid will take into account the total cost of the requirement inclusive of all duties, taxes, and other charges 16.5 The contract shall be awarded only to the bidder who is substantially technically responsive, offer the lowest rate. 16.6 The Purchaser shall compare all substantially responsive bids to determine the lowest evaluated bid.
F. AWARD OF CONTRACT	
17. Award Criteria	Pursuant to ITB Clause 17, the Purchaser will award the Contract to the Bidder whose bid has been determined to be substantially responsive and has been determined to be the lowest evaluated bid.
18. Purchaser's Right to Accept Any Bid and to Reject Any or All Bids	The Purchaser reserves the right to accept or reject any bid or to annul the bidding process and reject all bids at any time prior to contract award, without thereby incurring any liability to the affected Bidder or Bidders. No reason for such action of Purchaser shall be given.

SECTION II.

General Conditions of Contract/ Order

1. Packing	1.1 The Supplier shall provide such packing of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in the Contract.
2. Delivery and Documents	2.1 Documents to be submitted to the purchaser: - Upon the delivery of the Goods, the Supplier shall notify the Purchaser in writing and deliver to the Purchaser the following documents i. One original and one copy of the commercial invoice, indicating the DFY as the Purchaser, the Contract/order number, Goods' description, quantity, unit price, and total amount. Invoices must be signed in original and stamped or sealed with the company stamp/seal. ii. One original Acknowledgement of receipt of Goods/Final Acceptance by DFY.
3. Warranty	3.1 Bidder assigns and passes through to Purchaser the Manufacturer warranties on all "consumer products" as per company warranty term, wherever applicable. 3.3 Supplier shall be provided comprehensive maintenance warranty for all refurbishment, renovation work / Goods, supply for 1 year from the date of work completion. In case of any claim arising out of this warranty, the purchaser/consignee shall promptly notify the same in writing to the supplier. 3.4 Upon receipt of such notice, the supplier shall, with all reasonable speed repairs or replace the defective goods or parts thereof, free of cost, at the ultimate destination. The supplier shall take over the replaced parts/goods after providing their replacements and no claim whatsoever shall lie on the purchaser for such replaced parts/goods thereafter.
4. Payment	4.1 The method and conditions of payments to be made to the supplier shall be paid upon under this contract shall be mutually agreed upon and documented prior to contract award.



5. Prices	5.1 Prices charged by the Supplier for Goods delivered and Services performed under the Contract shall not vary from the prices quoted by the Supplier in its bid for the duration of the Contract. Prices shall be fixed and firm for the duration of the Contract. However, GST (as and when applicable) payable shall be paid as applicable at the time of supply.
6. Work/Goods Completion Timeline	6.1 The proposed Work/Goods must be completed within 45 days after issuing the contract/agreement or as per Purchase Order (PO)/Work Order (WO).
7. Liquidated Damages	If the Supplier fails to deliver any or all of the Goods or to perform the Services within the period(s) specified in the contract, the Purchaser shall, without prejudice to its other remedies under the Contract, deduct from the contract prices as liquidated damages, a sum equivalent to the 0.5 percent per week or part thereof of the delivered price of the delayed Goods or unperformed Services for each week or part thereof of delay until actual delivery or performance, up to a maximum deduction of the 10 percent of the value of delayed Goods. Once the maximum is reached, the Purchaser may consider termination of the contract.
8. Completion	8.1 The Supplier must notify the Purchaser's Representative in writing as soon as it considers that the Works have reached the stage of Substantial Completion. 8.2 Final Completion Certificate Supplier's obligations will not be considered to have been completed until the Purchaser's Representative has issued the Final completion Certificate to the Supplier, stating the date on which the Supplier completed its obligations under the Contract/Agreement. The Purchaser's Representative must issue the Final Completion Certificate within 10 days of the written notification of completion of work by the supplier.
9. Jurisdiction	All disputes arising out of the contract shall be subject to the jurisdiction of the appropriate court at New Delhi, India, only.

Section III

Technical Specifications

Health Centre Items (The delivery of these items is required 4 times on a same location)					
Sr. No	Name of item	Remarks	Units	Brand and Model	Detailed Specifications
General OPD					
1	B P Apparatus	Manual version	1	Brand Diamond manual BP apparatus	Type: Aneroid (dial) sphygmomanometer, manual, adult size. Dial: 60-75 mm, chrome/metal body, range 0-300 mmHg, graduation 2 mmHg, accuracy ± 3 mmHg. Cuff: Adult cotton/nylon cuff with Velcro, size 25-35 cm (arm), with latex-free inflation bulb and air-release valve (metal, easy control). Pressure-tight rubber tubing; carry pouch/box included. Standard: ISO 81060-1 / IS 3390 compliant;
2	Glucometer	1 machine with 100 Strip+ Lancet	1	SD Biosensor glucometer with SD Biosensor glucostrips	Type: Electronic blood glucose monitoring system (glucose oxidase / GDH-FAD biosensor strips), code-free. Measuring range: 20-600 mg/dL; sample: capillary whole blood, $\leq 1 \mu\text{L}$; result time ≈ 5 sec; memory ≥ 250 results with date/time and averages. Hematocrit range 20-60%; operating temp 10-40°C; auto on/off, battery CR2032 x1. Supply: 1 meter + 100 glucose test strips + 100 sterile lancets + lancing device + carry case + user manual. Strip expiry ≥ 12 months at delivery. CE / ISO 15197:2013 compliant, CDSCO registered.
3	Heightometer	Stature Meter 200 cm metal	1	Stainless steel	Stature meter (height measuring rod), portable/wall-mount, stainless steel/powder-coated metal. Range: 0-200 cm, graduation 1 mm (0.1 cm), accuracy ± 1 mm; sliding head-piece (horizontal arm) with locking. Detachable/telescopic sections with base plate; non-slip rubber feet; both cm & inch scale preferable. Rust-proof, durable; with carry bag.

4	Hb Meter	1 Machine with 50 Strip + lancet	1	Accusure with 25 Hb strips	Portable digital haemoglobin meter (photometric/biosensor strip method) with single-use test strips. Measuring range: 4.0-25.6 g/dL (Hct 15-65%); sample: capillary/venous whole blood $\approx$2-10 μL; result time $\approx$5-15 sec; accuracy $\pm$0.5-1.0 g/dL. Memory $\geq$100 results; auto strip detect; battery-operated. Supply: 1 meter + 50 Hb strips (Remarks ask 50; brand column mentions 25 - please confirm) + lancets + lancing device + control solution/code chip + manual. Strip expiry $\geq$12 months. CE / CDSCO registered
5	Baby weighing machine	Upto 20kg weight with scale	1	Accusure	Digital baby weighing scale, tray/pan type, for infants. Capacity: 0-20 kg, resolution 5-10 g (or 10 g); accuracy $\pm$10 g. LCD display, tare function, hold function, kg/lb selection, auto power-off, low-battery indicator. Stainless steel / ABS removable weighing tray; non-slip feet; battery or adapter powered. With calibration weight/certificate;
6	Adult weighing machine		1	Accsure	Digital adult weighing scale (platform/personal). Capacity: 0-180 kg (min 150 kg), resolution 100 g, accuracy $\pm$0.5 kg. Large LCD display ($\geq$25 mm digits), tare, auto on/off, overload & low-battery indicator. Tempered glass or ABS/steel platform with non-slip surface; battery operated (2xAAA / CR2032).
7	Stethoscope		1	Accusure or Dr. Morpean	Dual-head stethoscope, adult. Chestpiece: stainless steel/aluminium alloy, bell + diaphragm (reversible, $\approx$45 mm diaphragm); non-chill rim. Binaurals: chrome/anodised, spring-loaded with adjustable tension; soft silicone ear tips (latex-free). Tubing: PVC/latex-free Y-tubing, $\approx$28-30 inch ($\approx$56 cm), good acoustic sealing. With spare diaphragm and ear tips.
8	Torch		1	Cynor	Diagnostic torch (examination light), LED, pen/hand-held type. Bright white LED ($\geq$ 100 lumens), focused beam, ABS/aluminium body. Rechargeable (USB) or battery operated (2-3 AA/AAA), with on/off switch; battery & spare included. Suitable for throat/pupil exam.

9	vaccine carrier box	2.5 Lts	1	STOP Plastic vaccine carrier box	Vaccine carrier (cold box) – 2.5 L (as per Remarks), insulated with polyethylene/PU foam, plastic outer & inner shell. Supplied with 4 conditioned ice packs (as applicable to size); tight-fitting lid with foam pad, carrying strap/handle. Cold life $\geq$ 36-48 hrs at 43°C ambient (as per WHO PQS E004/VC); keeps +2°C to +8°C. Leak-proof, easy to clean; UIP-compliant.
10	Thermometer and pulse oximeter		1	Dr. Trust and Dr. Morpean	Combo: (a) Digital clinical thermometer – flexible/rigid tip, range 32-42.9°C, accuracy $\pm 0.1^\circ\text{C}$ (35.5-42°C), waterproof, memory last reading, beep alarm, auto off, LCD display; or infrared non-contact forehead thermometer (range 32-42.9°C, accuracy $\pm 0.2^\circ\text{C}$, < 1 sec). (b) Fingertip pulse oximeter – SpO ₂ 70-100% ($\pm 2\%$), pulse rate 30-250 bpm (± 2 bpm), OLED/LED display, auto power off, 2 x AAA batteries, lanyard & pouch. CE / CDSO
11	Pulse oximeter		1	Dr. Trust	Fingertip pulse oximeter, digital. SpO ₂ range 70-100% (accuracy $\pm 2\%$ at 70-100%); pulse rate 30-250 bpm (± 2 bpm). OLED/LED display with 4-direction rotation, plethysmograph/pulse bar, low-battery indicator, auto power-off (≤ 8 sec). Power: 2 x AAA; with lanyard, pouch, user manual. Suitable for adult & paediatric fingers. CE / ISO 80601-2-61;
12	Nebulizer	with 2 mask	1	Accusure	Compressor nebulizer, with 2 masks (1 adult + 1 paediatric) and mouthpiece. Power: 220-240 V AC, 50 Hz; compressor piston type; max pressure ≥ 30 psi (≈ 2 bar), airflow $\approx 6-8$ L/min. Nebulization rate ≥ 0.3 mL/min; MMAD $\approx 3-5$ μm ($\geq 70\%$ particles < 5 μm); medication cup $\approx 6-10$ mL; residual volume < 1 mL. Supplied with air tube, mask (adult/child), mouthpiece, spare filters; noise ≤ 60 dB. CE / CDSO;
13	Oxygen Concentrator	10 Lts with 2 mask	1	Oxymed/Nidek/inogen	Stationary oxygen concentrator, 10 L/min with 2 masks (adult + paediatric) and humidifier bottle. Flow: 0-10 L/min adjustable (flowmeter); O ₂ purity 93% $\pm 3\%$ at 10 L/min; outlet pressure 4-7 psi. Power: 230 V AC, 50 Hz, consumption $\approx 600-750$ W; noise ≤ 50 dB; continuous 24x7 operation. Alarms: power failure, low O ₂ purity, low pressure/flow, high temperature; hour-meter;

					wheels/casters; washable filter. Supply with 2 O ₂ masks, nasal cannula, tubing & humidifier. CE / ISO 8359;
14	ILR Freezer	50L	1	Daikin/Bluestar/Godrej	ILR / vaccine freezer, ~50 L net capacity (as per Remarks), top-opening. WHO-PQS prequalified (E003 category); storage +2°C to +8°C (ILR) with ≥ 48 hrs holdover at 43°C ambient. If a deep-freezer is intended (-15 to -25°C for ice-pack freezing), please confirm. Power: 220-240 V AC, 50 Hz; low-energy compressor; wide voltage tolerance (160-270 V). With digital temperature display/data-logger, alarm, lockable lid, basket(s), castors; refrigerant R600a/R134a (eco-friendly, ozone-friendly). Standard: WHO-PQS;
Essential Furniture			Units		
14	Examination Table	Adjustable Table with Step and Patient stool	1	Brand Coach	Hospital examination Table, with step stool and patient stool. Size ≈ 6 ft (L) x 2 ft (W) x 30 inch height; frame: 25 mm powder-coated SS 304 top: 2-section, Includes: foot-step (2-step), and a revolving patient stool. Load capacity ≥ 200 kg; non-slip rubber feet; rust-proof powder-coat finish.
15	Mattress for examination table	2x6 feet	1	Rexine	Examination table mattress – 2 x 6 ft (≈ 60 x 180 cm), thickness 2-3 inch (50-75 mm). High-density PU foam / rebonded foam (density ≥ 32 kg/m ³) with water-proof, fluid-resistant rexine (PVC leather) cover, easily wipeable; fire-retardant preferred. Stitched seam, non-slip bottom.
16	Office Table	Office Table with One Drawer and One cabinet	2	Brand Coach	Office table, 4 ft x 2 ft x 30 inch (approx. 120 x 60 x 76 cm). Top: 25 mm pre-laminated particle/MDF board (both sides laminate); frame powder-coated MS. With 1 drawer (lockable) and 1 side cabinet (lockable) - total 2 pieces. Colour as per tender (e.g. brown/grey).
17	Chair with arm	Steel Office Chair With Arm Rest	6	Brand Coach	Steel/metal office chair with armrest, fixed (non-revolving) type – 6 pcs. Frame: powder-coated MS tubular; seat & backrest: padded with PU foam (≈ 30-40 mm), covered with Rexine/fabric; moulded or foam armrests. Seat ≈ 18 x 18 inch; load ≥ 100 kg; rubber caps on legs.

18	Stool	Revolving	4	Brand Coach	Revolving stool (4 pcs), height-adjustable (gas lift / screw), 360° swivel. Seat: round, 12-14 inch (30-35 cm) dia, PU foam padded, rexine upholstery; 5-star / chrome-plated or powder-coated base with castors. Height adjustment range $\approx$ 18-24 inch; load $\geq$ 100 kg.
19	Office almirah		2	Brand Coach	Steel almirah (2 pcs), size $\approx$ 6 ft (H) x 3 ft (W) x 1.5 ft (D) (1800 x 900 x 450 mm). Made of 0.8-1.0 mm CRCA sheet steel (powder-coated); 2 hinged doors with lock; 4-5 adjustable shelves; reinforced. Levelling feet; rust-proof finish.
20	Instrument tray		2	Brand Coach	Instrument tray, stainless steel SS 304, with cover Size $\approx$ 12 x 8 x 2 inch (300 x 200 x 50 mm) or as per tender; 18 SWG; smooth, rolled edges, autoclavable, corrosion-resistant. Qty 2.
21	Medicine rack		3	Slot angle heavy duty rack	Medicine rack, slotted-angle heavy duty, MS powder-coated. Size $\approx$ 6 ft (H) x 3 ft (W) x 1.5 ft (D) (1830 x 915 x 460 mm), 5-6 shelves; angle thickness $\geq$ 2 mm; load $\geq$ 40-45 kg per shelf. With nut-bolt assembly; qty 3.
22	BMW - Medium	16 Lts (Blue, Yellow, Green)	2	Aristo or any reputed brand	BMW (bio-medical waste) bins, medium, 16 L each, colour-coded as per BMW Rules 2016: Blue, Yellow, Green (qty 2 sets). Material: virgin HDPE/PP, pedal-operated with lid, durable & leak-proof; non-chlorinated bags compatible; colour-coded with biohazard symbol where applicable.
23	Black Dust bin - Small	7 lts (Black)	4	Aristo or any reputed brand	Black dust bin, small, 7 L, black (general waste), pedal-operated / swing lid. Material: virgin HDPE/PP; qty 4.
24	Puncture proof container		2	Bio hazard or any reputed brand	Puncture-proof sharps container (white translucent, HDPE/PP), as per BMW Rules 2016 (Annex). Capacity $\approx$ 1.5-5 L; tamper-proof, leak-proof, with biohazard symbol and wide-mouth opening for needle/syringe disposal; sealable lid. Qty 2.
25	Needle Cutter		2	Bio hazard or any reputed brand	Needle/hub cutter (burner/destroyer), hand-held or wall-mounted. Hand-held type: plastic housing with stainless-steel/hardened cutter blade; cuts needle hub of syringes; collects cut needles in a container ($\geq$ 500 needles). Electric needle destroyer option (burner): 230 V AC; ISI/CE as available. Qty 2.

26	Rack for documents	60 x 24 x 15 Inch 6 Shelves - 45Kg load capacity	2	Slot angle heavy duty rack	Document rack, slotted-angle MS powder-coated. Size 60 x 24 x 15 inch (H x W x D) with 6 shelves; load capacity 45 kg per shelf (as per remarks); angle thickness ≥ 1.8 -2 mm, with nut-bolt assembly. Qty 2.
27	Office chair		2	Brand Coach	Office chair, revolving, mid-back, with armrest (2 pcs). PU foam padded seat/back, rexine/fabric; gas lift height adjustment, tilt/lock mechanism, 5-star nylon/ metal base with castors. Load ≥ 100 -120 kg; BIFMA compliant preferred.
28	Semi Fowler Bed		2	Brand Coach semi fowler bed with mattress with abs panel with collapsible railings with 3 inch wheel	Semi-Fowler bed (manual, 2-function: back-rest + knee-rest crank) – 2 pcs. Size ≈ 78 x 36 x 20 inch (L x W x H); frame: powder-coated MS/SS; bed top: perforated/ABS panels, head & foot ABS panels, with collapsible SS/ABS side railings (both sides) & 3-inch castors with brakes. Back-rest 0-75°, knee-rest 0-30°; with foam mattress (≥ 4 inch) rexine cover; I.V. rod holder; safe working load ≥ 150 -200 kg. Note- 16 gauge Pipe and 18 Gauge Sheet Supplied with mattress.
29	Bed Side Screen	4 Panels with Wheel	2	Brand Coach	Bed side screen, 4-panel folding with castors. Frame: powder-coated SS tubular (≈ 19 mm); panels: washable fabric/Rexine/ECG-curtain on frame, size per panel ≈ 2 ft (W) x 5-6 ft (H); wheel castors (≥ 2 inch). Foldable and portable; qty 2.
30	Computer table		1	Brand Coach	Computer table with keyboard tray, CPU stand/shelf and drawer. Size ≈ 4 ft x 2 ft x 30 inch; top: 25 mm pre-laminated board; powder-coated MS frame; slide-out keyboard tray; cable manager.
31	Computer Set	All in one computer with UPS/ Laptop with Bag	1	Dell/HP/Lenovo	Intel Core i5 Processor (13th Generation or higher), Minimum 16 GB RAM, 512 GB SSD, 15.6" Full HD Anti-Glare Display, HD/FHD Webcam, USB-C Port, Windows 11 Pro, Backpack and Wireless Mouse
32	Extension board		2	Anchor	Power extension board, 6-way (6A/16A) with individual switches and neon indicator. Flame-retardant ISI-marked body, overload protection (MCB/fuse), surge protector preferred; 3-pin, 3 m (min) heavy cable of 1.5 sq mm copper; rating 240 V / 6-16 A. Qty 2.

33	Printer	Black and White	1	HP or Cannon	Laser printer, monochrome (black & white) A4. Print speed $\geq$ 20-30 ppm; resolution 600 x 600 dpi (up to 1200); duplex (auto) preferred; USB 2.0 + Wi-Fi/Ethernet; monthly duty cycle $\geq$ 10,000 pages. With starter toner; 1-yr warranty (HP / Canon).
Day Care			Units		
34	Bed Sheet		2	Cotton size 6 x2	Bed sheet for single bed: 100% cotton, white/bleached, size 6 ft x 2 ft (as per list) - 2 pcs; plain weave, $\geq$ 40 TPI (or 120 GSM+), hemmed edges, colour-fast, washable.
35	Pillow		2	size 18 x 24	Pillow with cover: size 18 x 24 inch, hollow fibre/PU foam-filled, soft, washable cotton cover (white); qty 2.
36	Blanket	Single bed (Single Bed)	4	Any reputed brand	Blanket, single-bed, size $\approx$ 60 x 90 inch; soft fleece/woollen-acrylic, $\geq$ 300 GSM, colour-fast, machine-washable; qty 4.
37	Scissors	6 Inch SS	3	Alex/Medi	Surgical scissors, 6 inch ($\approx$ 15 cm), stainless steel (AISI 420 / 304), sharp-sharp or blunt/sharp, straight (or curved), with finger rings, autoclavable, corrosion-resistant, mirror finish; qty 3.
38	Chital Forcep	8 Inch	1	Alex/Medi	Cheattle (Chittle) sterilising forceps, 8 inch ($\approx$ 20 cm), stainless steel, with cheattle bowl/jar & lid (SS) if applicable; autoclavable; qty 1.
39	Dressing kit		2	Alex/Medi	Dressing kit (set): SS instruments – dissecting forceps (toothed & non-toothed), artery forceps, scissors, kidney tray, gallipot / bowls, instrument tray with lid. All stainless steel AISI 304/420, autoclavable. Qty 2.
40	Dressing Trolley		1	Alex/Medi	Dressing trolley, stainless steel SS 304, 2 shelves, size $\approx$ 24 x 16 x 32 inch; top with raised edges, bottom shelf; 4 swivel castors (2 locking); smooth welded finish; qty 1.
41	Fetoscope		2	Alex/Medi	Fetoscope (Pinard horn type), aluminium or wood/plastic; length $\approx$ 20 cm; with ear-piece / head-band optional; for auscultation of foetal heart sounds. Qty 2 (brand Alex/Medi).
42	Stethoscope		2	Accusure/Dr. Morpean	Dual-head stethoscope, adult, stainless steel chest-piece (bell/diaphragm), Y-type PVC tubing, adjustable binaurals and soft ear tips; spare diaphragm and ear tips (Dr. Morepen); qty 2.
43	Snellen chart		1	Rumax	Snellen chart (distance vision), standard 6-metre (20 ft); printed on durable card/wipeable plastic or laminated; sizes 6/60 to 6/6; letters + tumbling E; with matte finish; size $\approx$ 24 x 30 inch. Qty 1.



44	Sterilizer		1	Stainless steel	Instrument steriliser (boiler type), stainless steel SS 304, rectangular with lid and perforated tray; size $\approx 12 \times 6 \times 4$ inch, electric heating 1000-2000 W, 230 V; thermostat optional. Qty 1.
45	Autoclave	25 Ltr	1	Alex/Medi	Autoclave, vertical, 25 L, top-loading, electric, with SS 304 chamber. Chamber: SS 304; working pressure 15 psi (121°C) up to 20 psi; heating 2000-2500 W, 230 V; safety valve, pressure gauge, safety lock; with SS basket. Qty 1.
46	Autoclave drum		1	Alex/Medi	Autoclave drum (SS 304), round with perforated side-holes and sliding cover; for linen/dressing sterilisation; size $\approx 9 \times 9$ inch (or 12×12 inch); qty 1.
47	Water filter	Gravity water purifier	1	Havells	Gravity-based (non-electric) water purifier (Havells), capacity $\approx 7-10$ L storage; UF/ultra-filtration + activated carbon + sediment filter; food-grade ABS/plastic body, with tap; no electricity; removes bacteria, cysts, turbidity (e.g. Havells Digi / Gravity), 1-yr warranty.

Name_____

Place:_____ Signature of Bidder_____

Date:_____ Seal of the Bidder_____